

All the bidders who have participated in Tender No: 1.3C/APMSIDC/2026-27, Dated: 20.07.2026. (Bids opened on Dt:14.08.2026) for Procurement and supply of medical equipment for Up Gradation of PG Seats in existing Medical Colleges (2 years Rate Contract)., are hereby requested to send their clarifications, if any, regarding the remarks of Preliminary Technical Evaluation Report which is placed in website <https://apmsidc.ap.nic.in> or before 05.00 PM, Dt: 21.08.2026.

Mail ID: aphmhidc@gmail.com.

Sd/-

Executive Director
APMSIDC

Date: 19.08.2026

Mangalagiri, Guntur

As per the above Qualification Criteria, the tender documents (on-line documents submitted by the participated firm) are evaluated and the details are as follows,

Sl. No	Document Description	M/s. Colloq Marketing Pvt. Ltd.	Remarks
1	Process Fee:29,500/-	Online	Complies
2	EMD- OT Lights (Rs. 10,00,000)	Online Rs:10,00,000 ID:3270040	Complies
3	Bid Form Section VII-A 90 days	Page no:1	Complies
4	List of items offered with Make and Model details without prices.	Page no:53	Complies
5	Manufacturers Authorization, wherever required	Page no: 7	Complies
6	Past Performance Details Format B1 along with supporting documents	Page no:9-11 Supporting documents:	Not Complies
7	End-User Certificates or CA Certificate as per Format B2	Page no:13	Complies
8	Financial Capability Details Format B3 for Manufacturer	Page no: Not applicable	Complies
9	Financial Capability Details Format B3- A Distributor OT Lights:8,00,00,000	Page no: 15 Avg annual turnover: Rs.19.92Cr Net:9.57Cr	Complies
10	Details & proof of After-Sales Service facilities	Page no:17	Complies
11	Letter of authorization to sign the bids	Page no:19	Complies
12	Clause-by-clause commentary on technical specifications	Page no:23-31	Complies
13	Technical and Commercial deviations statements	Page no:21	Complies
14	Copy of the GST Certificate and	Page no: 231	Complies

Sl. No	Document Description	M/s. Colloq Marketing Pvt. Ltd.	Remarks
	Details of IT Returns- PAN / TIN copies.	33AAICS8026R1ZQ Page.no 246 AAICS8026R Page.no: 243,244,245 returns	
15	The Manufacturer, must have necessary quality certifications for both processes and products such as ISO 9001 (Quality Management System for Organization) and ISO 13485 (Quality Management System for Medical Devices).	Pg.No.37 ISO 13485:2016 Valid till:19.06.2028.	Complies
16	Full Quality Assurance System Approval Certificate Management System Certification for Medical Devices and their equivalent International Standards certificates (BIS/CE/USFDA etc) & AERB certificate.	Page no:33-34 USFDA Pg.No.41 EU Declaration of conformity	Complies
17	IEC 60601-1, IEC 60601-2, and IEC 60601-2-41 certificates for OT Lights	Not Submitted.	Not Complies
18	Memorandum of Articles	Not Submitted	Not Complies
19	All the uploaded Technical bid, to be attested by a Gazette Officer or properly notarized or self-attested	self-attested	Complies
20	General information about the bidder	Page no:51	Complies
21	Declaration Form	Page no:47	Complies
22	DPIIT Declaration (Annexure IX)	Page no:49	Complies

Past Performance Details:

Sl. No	Description	Qty	Qualifying Qty	Make & Model	Qty Supplied		
					2022-23	2023-24	2024-25
1	Operation Theatre Ceiling light double Dome	50	25	Dr.MACH & LED 6MC	--	---	

Remarks: Provisionally not qualified due to

- 1.Non-submission of Past Performance supporting documents
- 2.Non-submission of IEC 60601-1, IEC 60601-2, and IEC 60601-2-41 certificates for OT Lights
3. Non-Submission of Memorandum of Articles.

Sl. No	Document Description	M/s.Kendriya Bhandar	Remarks
1	Process Fee:29,500/-	Online	Complies
2	EMD- OT Lights (Rs. 10,00,000)	DD no:273100 Dt:11.08.2026.	Complies
3	Bid Form Section VII-A 90 days	Page no:261	Complies
4	List of items offered with Make and Model details without prices	Page no:266	Complies
5	Manufacturers Authorization, wherever required	Page no:267	Complies
6	Past Performance Details Format B1 along with supporting documents	Page no:297 Supporting documents:300-350	Complies
7	End-User Certificates or CA Certificate as per Format B2	Page no:351	Complies
8	Financial Capability Details Format B3 for Manufacturer	Page no: Not applicable	Complies
9	Financial Capability Details Format B3-A Distributor OT Lights:8.0 Cr	Page no:353 Avg annual turnover: Rs.5704.49Cr Net worth:206Cr	Complies
10	Details & proof of After-Sales Service facilities	Page no:405	Complies
11	Letter of authorization to sign the bids	Page no:407	Complies
12	Clause-by-clause commentary on technical specifications	Page no:415	Complies
13	Technical and Commercial deviations statements	Page no:417	Complies
14	Copy of the GST Certificate and Details of IT Returns- PAN / TIN copies.	Page no:01 PAN: AAHAC9531G Page no:3-7 GST :37 AAHAC9531G1ZM Page.no:47-260 IT Returns 22-23,23-24,24-25	Complies
15	The Manufacturer, must have necessary quality certifications for both processes and products such as ISO 9001 (Quality Management System for Organization) and ISO 13485 (Quality Management System for Medical Devices).	Pg. No.425 ISO 13485:2016 Valid till:07.11.2026	Complies
16	Full Quality Assurance System Approval Certificate Management System Certification for Medical Devices and their equivalent International Standards certificates (BIS/Notified CE/ European CE/USFDA etc) certificate.	Page no:433 USFDA	Complies
17	IEC 60601-1, IEC 60601-2, and IEC 60601-2-41 certificates for OT Lights	Not submitted	Not Complies

Sl. No	Document Description	M/s.Kendriya Bhandar	Remarks
18	Memorandum of Articles	Not Submitted	Complies
19	All the uploaded Technical bid, to be attested by a Gazette Officer or properly notarized or self-attested	self-attested	Complies
20	General information about the bidder	Page no:449	Complies
21	Declaration Form	Page no:451	Complies
22	DPIIT Declaration (Annexure IX)	Page no:453	Complies

Past Performance Details:

Sl. No	Description	Qty	Qualifying Qty	Make & Model	Qty Supplied		
					2022-23	2023-24	2024-25
1	Operation Theatre Ceiling light double Dome	50	25	Hospedia & LED 50 Double Dome Sony Camera Model: FCB-EV 7520		16	56

Remarks: Provisionally not qualified due to Non-submission of IEC 60601-1, IEC 60601-2, and IEC 60601-2-41 certificates.

Sl. No	Document Description	M/s. Sai Srinivasa Agencies	Remarks
1	Process Fee:29,500/-	Online	Complies
2	EMD- OT Lights (Rs. 10,00,000)	BG: 0008NDLG00084227 Rs:10,00,000 ICICI bank	Complies
3	Bid Form Section VII-A 90 days	Page no:9	Complies
4	List of items offered with Make and Model details without prices .	Page no:10	Complies
5	Manufacturers Authorization, wherever required	Page no: 11	Complies
6	Past Performance Details Format B1 along with supporting documents	Page no:13-14 Supporting documents:15-62	Complies
7	End-User Certificates or CA Certificate as per Format B2	Page no:63	Complies
8	Financial Capability Details Format B3 for Manufacturer	Page no: Not applicable	Complies
9	Financial Capability Details Format B3-A Distributor OT Lights:8,00,00,000	Page no: 65 Avg annual turnover: Rs.82.86Cr Net:90.84Cr	Complies

Sl. No	Document Description	M/s. Sai Srinivasa Agencies	Remarks
10	Details & proof of After-Sales Service facilities	Page no:67	Complies
11	Letter of authorization to sign the bids	Page no:69	Complies
12	Clause-by-clause commentary on technical specifications	Page no:70-72	Complies
13	Technical and Commercial deviations statements	Page no:75	Complies
14	Copy of the GST Certificate and Details of IT Returns- PAN / TIN copies.	Page no: 76 36ACWFS9933Q1ZO Page.no 77 ACWFS9933Q Page.no: 80,83,86 returns	Complies
15	The Manufacturer, must have necessary quality certifications for both processes and products such as ISO 9001 (Quality Management System for Organization) and ISO 13485 (Quality Management System for Medical Devices).	Pg.No.109 ISO 13485:2016 Valid till:02.04.2029.	Complies
16	Full Quality Assurance System Approval Certificate Management System Certification for Medical Devices and their equivalent International Standards certificates (BIS/CE/USFDA etc) & AERB certificate.	Page no:112 EU Declaration of conformity Pg.No.119 USFDA	Complies
17	IEC 60601-1, IEC 60601-2, and IEC 60601-2-41 certificates for OT Lights	Pg.No. 120 - 135	Complies
18	Memorandum of Articles	NA	Complies
19	All the uploaded Technical bid, to be attested by a Gazette Officer or properly notarized or self-attested	self-attested	Complies
20	General information about the bidder	Page no:136	Complies
21	Declaration Form	Page no:137	Complies
22	DPIIT Declaration (Annexure IX)	Page no:138	Complies

Past Performance Details:

Sl. No	Description	Qty	Qualifying Qty	Make & Model	Qty Supplied		
					2022-23	2023-24	2024-25
1	Operation Theatre Ceiling light double Dome	50	25	Xcellance medical Technologies (Shalya) & ILUX 40-40	--	28	48

Remarks: Provisionally Qualified

Sl. No	Document Description	M/s.SMPL Life science Pvt.Ltd	Remarks
1	Process Fee:29,500/-	Online	Complies
2	EMD- OT Lights (Rs. 10,00,000)	Submitted document is blank	Not Complies
3	Bid Form Section VII-A 60 days	Submitted	Complies
4	List of items offered with Make and Model details without prices	Submitted	Complies
5	Manufacturers Authorization, wherever required	Submitted	Complies
6	Past Performance Details Format B1 along with supporting documents	Submitted	Complies
7	End-User Certificates or CA Certificate as per Format B2	Submitted	Complies
8	Financial Capability Details Format B3 for Manufacturer	Page no: Not applicable	Complies
9	Financial Capability Details Format B3- A Distributor OT Lights:8,00,00,000	Avg annual turnover: Rs.165.44Cr Net worth:31.28Cr	Complies
10	Details & proof of After-Sales Service facilities	Submitted	Complies
11	Letter of authorization to sign the bids	Submitted	Complies
12	Clause-by-clause commentary on technical specifications	Submitted	Complies
13	Technical and Commercial deviations statements	Submitted	Complies
14	Copy of the GST Certificate and Details of IT Returns- PAN / TIN copies.	PAN No: AAQCS7546E GST:06AAQCS7546E1ZZ IT Returns :Submitted	Complies
15	The Manufacturer, must have necessary quality certifications for both processes and products such as ISO 9001 (Quality Management System for Organization) and ISO 13485 (Quality Management System for Medical Devices).	Pg. No. ISO 13485:2016 Valid till:22.03.2029	Complies
16	Full Quality Assurance System Approval Certificate Management System Certification for Medical Devices and their equivalent International Standards certificates (BIS/Notified CE/European CE/USFDA etc) certificate.	USFDA (Surgident India)- Submitted	Complies
17	IEC 60601-1, IEC 60601-2, and IEC 60601-2-41 certificates for OT Lights	Not submitted	Not Complies

Sl. No	Document Description	M/s.SMPL Life science Pvt.Ltd	Remarks
18	Memorandum of Articles	Submitted	Complies
19	All the uploaded Technical bid, to be attested by a Gazette Officer or properly notarized or self-attested	self-attested	Complies
20	General information about the bidder	Submitted	Complies
21	Declaration Form	Submitted	Complies
22	DPIIT Declaration (Annexure IX)	Submitted	Complies

Past Performance Details:

Sl. No	Description	Qty	Qualifying Qty	Make & Model	Qty Supplied		
					2022-23	2023-24	2024-25
1	Operation Theatre Ceiling light double Dome	50	25	Surgident India & GAMA 18+ 18			299

Remarks: Provisionally not qualified due to

- Non submission of hard copy of EMD.
- Non-Submission of IEC 60601-1, IEC 60601-2, and IEC 60601-2-41 certificates for OT Lights.

Sl. No	Document Description	M/s. Unifranco Life Sciences Pvt Ltd	Remarks
1	Process Fee:29,500/-	Online	Complies
2	EMD- OT Lights (Rs. 10,00,000)	DD No. 503710 Rs:10,00,000 DD Dt: 14.08.2026.	Not Complies
3	Bid Form Section VII-A 90 days	Submitted	Complies
4	List of items offered with Make and Model details without prices	Submitted	Complies
5	Manufacturers Authorization, wherever required	Submitted	Complies
6	Past Performance Details Format B1 along with supporting documents	Non-Submission of PO and satisfactory reports	Not Complies
7	End-User Certificates or CA Certificate as per Format B2	Submitted	Complies
8	Financial Capability Details Format B3 for Manufacturer	Page no: Not applicable	Complies
9	Financial Capability Details Format B3-A Distributor	Avg annual turnover: Rs.12.27Cr Net worth:5.28Cr	Complies

Sl. No	Document Description	M/s. Unifranco Life Sciences Pvt Ltd	Remarks
	OT Lights:8,00,00,000		
10	Details & proof of After-Sales Service facilities	Submitted	Complies
11	Letter of authorization to sign the bids	Submitted	Complies
12	Clause-by-clause commentary on technical specifications	Submitted	Complies
13	Technical and Commercial deviations statements	Submitted	Complies
14	Copy of the GST Certificate and Details of IT Returns- PAN / TIN copies.	PAN: AAACU8109P GST:36AAACU8109P1ZU IT Return:	Complies
15	The Manufacturer, must have necessary quality certifications for both processes and products such as ISO 9001 (Quality Management System for Organization) and ISO 13485 (Quality Management System for Medical Devices).	ISO 13485:2016 Valid till:02.08.2029	Complies
16	Full Quality Assurance System Approval Certificate Management System Certification for Medical Devices and their equivalent International Standards certificates (BIS/notified CE/European CE/USFDA etc) certificate.	Not Submitted	Not Complies
17	IEC 60601-1, IEC 60601-2, and IEC 60601-2-41 certificates for OT Lights	Submitted.	Not Complies
18	Memorandum of Articles	Not Submitted	Not Complies
19	All the uploaded Technical bid, to be attested by a Gazette Officer or properly notarized or self-attested	self-attested	Complies
20	General information about the bidder	Submitted	Complies
21	Declaration Form	Submitted	Complies
22	DPIIT Declaration (Annexure IX)	Not Submitted	Not Complies

Past Performance Details:

Sl. No	Description	Qty	Qualifying Qty	Make & Model	Qty Supplied		
					2022-23	2023-24	2024-25
1	Operation Theatre Ceiling light double Dome	50	25	Lumens Medical Pvt Ltd & Blitz L6 Twin Dome	--	---	

Remarks: Provisionally not qualified due to

1. Non submission of hard copy of EMD.
2. Non-submission of Purchase orders and satisfactory reports for quoted item.
3. Non-submission of BIS/notified CE/European CE/USFDA etc.
4. Non-submission of DPIIT Declaration (Annexure IX).

Sl. No	Document Description	M/s.BR INFRACON	Remarks
1	Process Fee:29,500/-	Online	Complies
2	EMD- OT Lights (Rs. 10,00,000)	Online ID:3271964 Payment dt: 13.08.2026	Complies
3	Bid Form Section VII-A 90 days	Page no:3	Complies
4	List of items offered with Make and Model details without prices	Page no:4	Complies
5	Manufacturers Authorization, wherever required	Page no: 5	Complies
6	Past Performance Details Format B1 along with supporting documents	Page no:6-6(a) Supporting documents:6(b)-6(Z)	Complies
7	End-User Certificates or CA Certificate as per Format B2	Page no:7	Complies
8	Financial Capability Details Format B3 for Manufacturer	Page no: Not applicable	Complies
9	Financial Capability Details Format B3-A Distributor OT Lights:8,00,00,000	Page no: 9 Avg annual turnover: Rs.40 Cr Net:40 Cr	Complies
10	Details & proof of After-Sales Service facilities	Page no:10	Complies
11	Letter of authorization to sign the bids	Page no:11	Complies
12	Clause-by-clause commentary on technical specifications	Page no:12-12b	Complies
13	Technical and Commercial deviations statements	Page no:13	Complies
14	Copy of the GST Certificate and Details of IT Returns- PAN / TIN copies.	Page no: 14 36AARFB6050C2Z1 Page.no 14A AARFB6050 Page.no: 14B-14F IT returns	Complies
15	The Manufacturer, must have necessary quality certifications for both processes and products such as ISO 9001 (Quality Management System for Organization) and ISO 13485 (Quality Management System for Medical Devices).	Page no:15 ISO 13485:2016 Valid till:06.09.2028	Complies
16	Full Quality Assurance System Approval Certificate Management System Certification for Medical Devices and their equivalent International Standards certificates	USFDA Certificate	Complies

Sl. No	Document Description	M/s.BR INFRACON	Remarks
	(BIS/CE/USFDA etc) & AERB certificate.		
17	IEC 60601-1, IEC 60601-2, and IEC 60601-2-41 certificates for OT Lights	Not submitted	Not Complies
18	Memorandum of Articles	Not Applicable	Complies
19	All the uploaded Technical bid, to be attested by a Gazette Officer or properly notarized or self-attested	self-attested	Complies
20	General information about the bidder	Page no:19	Complies
21	Declaration Form	Page no:20	Complies
22	DPIIT Declaration (Annexure IX)	Page no:21	Complies

Past Performance Details:

Sl. No	Description	Qty	Qualifying Qty	Make & Model	Qty Supplied		
					2022-23	2023-24	2024-25
1	Operation Theatre Ceiling light double Dome	50	25	Cognate & Vision 18+18	123	20	1

Remarks: Provisionally not qualified due to non-submission of IEC 60601-1, IEC 60601-2, and IEC 60601-2-41 certificates for OT Lights.

Sl. No	Document Description	M/s.Stann Bio Med Engineering Pvt. Ltd.	Remarks
1	Process Fee:29,500/-	Online	Complies
2	EMD- OT Lights (Rs. 10,00,000)	BG:00359231BG000294 Dt:14.08.2026 Indian Bank Rs:10,00,000	Complies
3	Bid Form Section VII-A 90 days	Page no:3	Complies
4	List of items offered with Make and Model details without prices.	Page no:4	Complies
5	Manufacturers Authorization, wherever required	Page no: 5	Complies
6	Past Performance Details Format B1 along with supporting documents	Page no:6-13 Supporting documents:13-101	Complies
7	End-User Certificates or CA Certificate as per Format B2	Page no:102	Complies
8	Financial Capability Details Format B3 for Manufacturer	Page no: Not applicable	Complies
9	Financial Capability Details Format B3-A Distributor	Page no: 103 Avg annual turnover: Rs.65.93Cr Net:16.6 Cr	Complies

Sl. No	Document Description	M/s.Stann Bio Med Engineering Pvt. Ltd.	Remarks
	OT Lights:8,00,00,000		
10	Details & proof of After-Sales Service facilities	Page no:104	Complies
11	Letter of authorization to sign the bids	Page no:105	Complies
12	Clause-by-clause commentary on technical specifications	Page no:106-108	Complies
13	Technical and Commercial deviations statements	Page no:230	Complies
14	Copy of the GST Certificate and Details of IT Returns- PAN / TIN copies.	Page no: 231 33AAICS8026R1ZQ Page.no 246 AAICS8026R Page.no: 243,244,245 returns	Complies
15	The Manufacturer, must have necessary quality certifications for both processes and products such as ISO 9001 (Quality Management System for Organization) and ISO 13485 (Quality Management System for Medical Devices).	Page no:247 ISO 9001:2015 Valid Upto:18.01.2028 Pg.No.248 ISO 13485:2016 Valid till:27.06.2029	Complies
16	Full Quality Assurance System Approval Certificate Management System Certification for Medical Devices and their equivalent International Standards certificates (BIS/CE/USFDA etc) & AERB certificate.	Page no:249-251 USFDA Pg.No.252 Verification of conformity Valid up to:13.02.2030.	Complies
17	IEC 60601-1, IEC 60601-2, and IEC 60601-2-41 certificates for OT Lights	Pg.No.112-191	Complies
18	Memorandum of Articles	Not Submitted	Not Complies
19	All the uploaded Technical bid, to be attested by a Gazette Officer or properly notarized or self-attested	self-attested	Complies
20	General information about the bidder	Page no:255	Complies
21	Declaration Form	Page no:257	Complies
22	DPIIT Declaration (Annexure IX)	Page no:258	Complies

Past Performance Details:

Sl. No	Description	Qty	Qualifying Qty	Make & Model	Qty Supplied		
					2022-23	2023-24	2024-25

1	Operation Theatre Ceiling light double Dome	50	25	STANN Bio-Med Engineering Pvt. Ltd. STANN / Discover	--	---	56
---	---	----	----	--	----	-----	----

Remarks: Provisionally not qualified due to Non-Submission of Memorandum of Articles.

Sl. No	Document Description	M/s. ZENITH MEDICAL DEVICES	Remarks
1	Process Fee:29,500/-	Online	Complies
2	EMD- OT Lights (Rs. 10,00,000)		Complies
3	Bid Form Section VII-A 9s0 days	Page no:	Complies
4	List of items offered with Make and Model details without prices (Dental X-ray & Laptop)	Page no:	Complies
5	Manufacturers Authorization, wherever required	Page no:	Complies
6	Past Performance Details Format B1 along with supporting documents	Page no: Supporting documents:	Not Complies
7	End-User Certificates or CA Certificate as per Format B2	Page no:	Complies
8	Financial Capability Details Format B3 for Manufacturer	Page no: Not applicable	Complies
9	Financial Capability Details Format B3- A Distributor OT Lights:8,00,00,000	Page no: Avg annual turnover: Rs.Cr Net worth:Cr	Complies
10	Details & proof of After-Sales Service facilities	Page no:	Complies
11	Letter of authorization to sign the bids	Page no:	Complies
12	Clause-by-clause commentary on technical specifications	Page no:	Complies
13	Technical and Commercial deviations statements	Page no:	Complies
14	Copy of the GST Certificate and Details of IT Returns- PAN / TIN copies.	Page no: PAN Page.no Page.no:	Complies
15	The Manufacturer, must have necessary quality certifications for both processes and products such as ISO 9001 (Quality Management System for Organization) and ISO 13485 (Quality Management System	Pg. No. ISO 13485:2016 Valid till:	Complies

Sl. No	Document Description	M/s. ZENITH MEDICAL DEVICES	Remarks
	for Medical Devices).		
16	Full Quality Assurance System Approval Certificate Management System Certification for Medical Devices and their equivalent International Standards certificates (BIS/CE/USFDA etc) & AERB certificate.	Page no: Pg.No.	Complies
17	IEC 60601-1, IEC 60601-2, and IEC 60601-2-41 certificates for OT Lights	Pg.No. Not Submitted.	Not Complies
18	Memorandum of Articles	Not Submitted	Not Complies
19	All the uploaded Technical bid, to be attested by a Gazette Officer or properly notarized or self-attested	self-attested	Complies
20	General information about the bidder	Page no:	Complies
21	Declaration Form	Page no:	Complies
22	DPIIT Declaration (Annexure IX)	Page no:	Complies

Past Performance Details:

Sl. No	Description	Qty	Qualifying Qty	Make & Model	Qty Supplied		
					2022-23	2023-24	2024-25
1	Operation Theatre Ceiling light double Dome	50	25	Staan & Discovery	--	---	

Remarks: Provisionally not qualified, as the same make and model has been quoted directly by the OEM, M/s. Staas. Hence, the bid submitted by your firm was not evaluated further and is treated as disqualified.